



# TEST CERTIFICATE

ISSUED BY SIRA TEST & CERTIFICATION LIMITED

## TEST FOR THE INGRESS PROTECTION OF AN nVISION HAND HELD CALIBRATOR

**Supplier:** Crystal Engineering Corporation  
708 Fiero Lane  
Suite 9  
San Luis Obispo  
California 93401  
USA

**Model or Type Identification:** nVision hand held calibrator

**Standard:** BS EN 60529:1992 Incorporating Amendments Nos 1 and 2

**Deviations from Standard:** None

**ST&C Test Procedure:** LOP 220

**ST&C Test Reports:** 09/0179 and N52A19587A

**Samples Delivery Date:** 9 March 2009

**Tests Conducted Between:** 23 to 26 March 2009

This certificate refers to the performance of the test samples when tested against the agreed programme. It does not imply that any other samples or products necessarily comply with the requirements of the test programme.

Sira Test & Certification Limited being a UKAS accredited Test House in accordance with ISO/IEC 17025 has tested the above nVision hand held calibrator, and has found it to comply with the requirements of the Ingress Protection Code: IP 67.

**S P Cork**  
**Laboratory Manager**

**Dated** 14 April 2009

*This certificate may only be reproduced in its entirety and without change*

**Page 1 of 1**

**Certificate No: Sira 52A19587ALab**  
**Form 6240, Issue 8**

a Volvere plc group company