



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1450352 (110579)

Master Contract: 187869

Project: 1585088

Date Issued: 2004/08/10

Issued to: Crystal Engineering Corp.
720 Aerovista Pl Suite B
San Luis Obispo, California 93401
USA
Attention: David K. Porter

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Pat Conway

Authorized by: Patricia Pasemko, Operations Manager

PRODUCTS

CLASS 2258 03 - PROCESS CONTROL EQUIPMENT - Intrinsically Safe and Non Incendive Systems - For Hazardous

CLASS 2258 83 - PROCESS CONTROL EQUIPMENT-Intrinsically Safe and Non-Incendive - Systems-For Hazardous Locations-Certified to U.S. Standards

Class I, Division 1, Groups A, B, C, D

Series XP2i and XP2i-DD digital test gauges; Temperature Code T4; Maximum Working Pressure 10,000 PSI; Intrinsically Safe when powered by one of the following batteries (3-cells, AA, max 4.95V); Rayovac

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1450352 (110579)

Master Contract: 187869

Project: 1585088

Date Issued: 2004/08/10

Maximum Plus, Eveready E91, Eveready EN91, Duracell MN1500.

Note: Prefixes and Suffixes are added to denote pressure range and options.

APPLICABLE REQUIREMENTS

C22.2 No 0 - M1991 General Requirements - Canadian Electrical Code Part II.

C22.2 No 157 - M1992 Intrinsically Safe and Non-Incendive Equipment for Use in Hazardous Locations.

UL 913, Sixth Edition Intrinsically Safe Apparatus and Associated Apparatus for use in Class I, II, III, Division 1, Hazardous (Classified) Locations.